



Sample Extraction

1. Label a soil extraction bottle and amber glass vial with the sample identification.
2. Weigh 5 g of the soil sample using a weigh boat and wooden spatula.
3. Transfer the sample to the soil extraction bottle.
4. Measure 10 mL of methanol and add it to the soil extraction bottle. The disposable plastic beaker can be used as an intermediate container for the solvent.
5. Cap the soil extraction bottle tightly and shake vigorously for a minimum of 2 minutes. Exposure of the sample to the solvent longer than two minutes is not detrimental.
6. Allow the sample to settle for 2-3 minutes before proceeding.
7. Transfer a portion of the liquid extract to the base of the filtration device.
8. Insert the plunger portion of the filtration device into the base and gently apply pressure to force the extract through the filter.
9. Transfer the filtrate to the appropriate amber glass vial using a disposable transfer pipette and cap tightly.
10. The filtrate can now be analyzed using the Beacon Toxaphene Tube Kit.

Sample Storage

Storage of soil samples should follow the holding conditions recommended for EPA methods 8100 and 8310 (GC and HPLC analysis of Toxaphene in soil, respectively) which is up to 14 days at 4°C (39°F).

Technical Assistance

For questions regarding this kit or for additional information about Beacon products, contact us.

Beacon Analytical Systems, Inc.
82 Industrial Park Rd. Saco, ME 04046
Tel. 207-571-4302
info@beaconkits.com | www.beaconkits.com

Safety

Stop Solution is 1N hydrochloric acid. Handle with care. To receive complete safety information on this product, contact Beacon Analytical Systems, Inc., and request Safety Data Sheets.

General Limited Warranty

Beacon Analytical Systems, Inc. ("Beacon") warrants the products manufactured by it against defects in materials and workmanship when used in accordance with the applicable instructions for a period not to extend beyond a product's printed expiration date. BEACON MAKES NO OTHER WARRANTY, EXPRESSED OR IMPLIED. THERE IS NO WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. The warranty provided herein and the data, specifications and descriptions of Beacon products appearing in published catalogues and product literature may not be altered except by express written agreement signed by an officer of Beacon. Representations, oral or written, which are inconsistent with this warranty, or such publications are not authorized and, if given, should not be relied upon. In the event of a breach of the foregoing warranty, Beacon's sole obligation shall be to repair or replace, at its option, any product or part thereof that proves defective in materials or workmanship within the warranty period, provided the customer notifies Beacon promptly of any such defect. The exclusive remedy provided herein shall not be deemed to have failed in its essential purpose so long as Beacon is willing and able to repair or replace any nonconforming Beacon product or part. Beacon shall not be liable for consequential, incidental, special or any other indirect damages resulting from economic loss or property damage sustained by a customer from the use of its products. However, in some states the purchaser may have rights under state law in addition to those provided by this warranty.

Intended Use

The Beacon Soil Extraction Kit is used to extract soil samples for testing in the Beacon Toxaphene Tube Kits.

Reagents and Materials Provided

- 100 Soil extraction bottles containing mixing beads and sodium sulfate.
- 100 Whatman uniprep two-piece filtration units.
- 100 Weigh boats.
- 100 Wood spatulas.
- 100 Amber glass vials with caps.
- 100 Disposable transfer pipettes.
- 1 Disposable beaker.

Reagents and Materials Required but Not Provided

- Pipette(s) with disposable tips capable of measuring 10 mL of methanol.
- Balance capable of weighing 5 g of sample.
- Methanol.
- Timer.
- Permanent marker.
- Personal protective equipment.

Kit Handling Notes and Precautions

- Read the product brochure in its entirety prior to use.
- All procedural steps should be completed without interruption. Ensure all reagents, materials and equipment are ready at the appropriate time.
- Dilution or adulteration of reagents or samples not called for in the procedure may result in inaccurate results.
- Use clean materials for each sample to avoid cross contamination.

Sample Collection

Distribution of Toxaphene in soil samples may be highly variable. This variability can be minimized through the use of composite sampling technique.

Sample Preparation

1. Remove any twigs, organic matter, and rocks.
2. Dry soil samples expected to contain > 30% water by weight. This includes but is not limited to samples taken from areas adjacent to standing water and surface soils collected during or immediately after rain or snow. Contact technical support for recommended drying methods.
3. Homogenize the soil sample.